Annexure Specification Particle Counter		To Be Filled by vendor
Range (Adjustable)	1.0 - 60,000 cumulative count / ml	
Software	User friendly Software	
Operating temperature	Ambient 0 to 70 °C	
Sample Nature	Fuel (Jet Fuel), Lube Oil and Water (Industrial Water, Effulent Water)	
Accessories & Requirements	All required accessories must be supplied	
Standard Accessories	All standard accessories must be included in the offer	
Consumables	A kit of basic consumables for two years trouble free operation must be included in the offer	\(\frac{1}{2}\)
Calibration & Refrence Standard	Certified & Refrence standards specially designed and produced for on-site qualification of equipment during installation and must be as per method	
nstallation & Commisioning	Must be installed & commisioned by manufacturer's representative	
raining	A comprehensive On-site training from Trained manufacture representative to our staff is required	
Required Documents	Operation Manual	
	Service Manual (including all Electrical, Electronics, & Mechanical schematics)	
	Factory Calibration Certificate	
	Factory Test Certificate	
	Certificate of Origin	

Kan sho